



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Non-Computing Scale
Digital Electronic
Model: 420 Series
 n_{max} : see table below
 e_{min} : see table below
Capacity: see table below
Platform: see table below
Accuracy Class: II

Submitted By:

Cardinal Scale Manufacturing Company
102 East Daugherty Street
Webb City, MO 64870
Tel: 417-673-4631 x 212
Fax: 417-673-2153
Contact: Thomas Schuller
Email: tschuller@cardet.com
Website: www.cardet.com

Standard Features and Options

- Semi-Automatic Zero Setting Mechanism (Push Button)
- Initial Zero Setting Mechanism (IZSM)
- Automatic Zero Tracking (AZT)
- Semi-Automatic Tare (Push Button)
- Alphanumeric Display
- AC Power
- DC Power
- Sleep Mode
- Auto Shut Off
- Single Range
- RS-232/USB

Model	Capacity	e_{min}	d	n_{max}	Platform	Class
420-300	300 g	0.05 g	0.005 g	6000	118 mm diameter	II
420-600	600 g	0.1 g	0.01 g	6000	145 mm x 115 mm	II
420-1200	1200 g	0.2 g	0.02 g	6000	145 mm x 115 mm	II
420-1500	1500 g	0.2 g	0.02 g	7500	145 mm x 115 mm	II
420-3000	3000 g	0.5 g	0.05 g	6000	188 mm x 148 mm	II

Load Cell Used: HBM Model SPL and HBM model PW6C (non-NTEP)

Temperature Range: 5 °C to 40 °C (37 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices. Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. *Editorial changes, not affecting the type or metrological content, corrected this certificate.

Kevin Schnepf
Chair, NCWM, Inc.

Marc Paquette
Chair, NTEP Committee
Issued: March 8, 2026

9011 South 83rd Street | Lincoln, Nebraska 68516

The National Council on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Cardinal Scale Manufacturing Company, Inc.
Non-Computing Scale / 420 Series

Application: For use in general purpose weighing applications.

Identification: A self-destructive sticker with the identification information is located on the side of the unit.

Sealing: A paper security seal is placed over a hole on the underside of the unit which covers the calibration switch.

Note: To verify that the scale is not in calibration mode, hold the calibration button (CAL) for 3 seconds.

Test Conditions: This certificate supersedes Certificate of Conformance Number 20-048 and was issued without additional testing to make editorial changes and clarify the test conditions. Previous test conditions are listed below for reference.

Certificate of Conformance 20-049: The emphasis of this evaluation was on the design, marking requirements, operation and compliance with influence factor requirements. Three scale capacities were provided for this evaluation: a 300g, a 1500g and a 3000g. These devices were tested over a temperature range of 5 °C to 40 °C (37 °F to 104 °F). They also had a load of approximately one-half capacity applied over 100,000 times with several increasing / decreasing load tests performed during this time. Several shift and load discrimination tests were conducted during the evaluation. The units under evaluation were tested with supply voltages from 102 to 132 VAC, 4.2 to 6.6 VDC and were also subjected to RFI testing.

Evaluated By: E. Morabito (NY); G. Gholston (NCWM) 20-049A1 (CN 11714)

Type Evaluation Criteria Used: *NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices*, 2016 Edition. *NCWM Publication 14 Measuring Devices*, 2016 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: D. Flocken (NCWM); J. Gibson (NCWM) 20-049A1

Examples of Device:



Unit shown without optional top enclosure



Calibration access sealing via self-destructive sticker