



**Med Vue**®



**Model MV2  
Medical Weight Analyzer  
Owner's Manual**



# INTRODUCTION

Thank you for purchasing our MedVue® Model MV2 Medical Weight Analyzer. It has been manufactured with quality and reliability and has been tested before leaving our factory to insure accuracy and dependability for years to come.

This manual is provided to guide you through installation, operation and maintenance of the MedVue Model MV2 Medical Weight Analyzer. Please read it thoroughly before attempting to install or operate the MV2 and keep it handy for future reference.




# COPYRIGHT

All rights reserved. Reproduction or use, without expressed written permission, of editorial or pictorial content, in any manner, is prohibited. No patent liability is assumed with respect to the use of the information contained herein.

# DISCLAIMER

While every precaution has been taken in the preparation of this manual, the Seller assumes no responsibility for errors or omissions. Neither is any liability assumed for damages resulting from use of the information contained herein. All instructions and diagrams have been checked for accuracy and ease of application; however, success and safety in working with tools depend to a great extent upon the individual accuracy, skill and caution. For this reason the Seller is not able to guarantee the result of any procedure contained herein. Nor can they assume responsibility for any damage to property or injury to persons occasioned from the procedures. Persons engaging the procedures do so entirely at their own risk.

<b>Serial Number</b> _____
<b>Date of Purchase</b> _____
<b>Purchased From</b> _____
_____
_____
_____
RETAIN THIS INFORMATION FOR FUTURE USE

<b>PRECAUTIONS</b>					
<b>Before using the weight analyzer, read this manual and pay special attention to all "NOTIFICATION" symbols:</b>					
	<b>IMPORTANT</b>		<b>ELECTRICAL WARNING</b>		<b>STATIC SENSITIVE</b>

## FCC COMPLIANCE STATEMENT


This equipment generates uses and can radiate radio frequency and if not installed and used in accordance with the instruction manual, may cause interference to radio communications. It has been tested and found to comply with the limits for a Class A computing device pursuant to Subpart J of Part 15 of FCC rules, which are designed to provide reasonable protection against such interference when operated in a commercial environment. Operation of this equipment in a residential area may cause interference in which case the user will be responsible to take whatever measures necessary to correct the interference.

You may find the booklet "How to Identify and Resolve Radio TV Interference Problems" prepared by the Federal Communications Commission helpful. It is available from the U.S. Government Printing Office, Washington, D.C. 20402. Request stock No. 001-000-00315-4.

## PROPER DISPOSAL

When this device reaches the end of its useful life, it must be properly disposed of. It must not be disposed of as unsorted municipal waste. Within the European Union, this device should be returned to the distributor from where it was purchased for proper disposal. This is in accordance with EU Directive 2002/96/EC. Within North America, the device should be disposed of in accordance with the local laws regarding the disposal of waste electrical and electronic equipment.

It is everyone's responsibility to help maintain the environment and to reduce the effects of hazardous substances contained in electrical and electronic equipment on human health. Please do your part by making certain that this device is properly disposed of. The symbol shown to the right indicates that this device must not be disposed of in unsorted municipal waste programs.



## CAUTION



**CAUTION:** RISK OF EXPLOSION IF BATTERY IS REPLACED BY AN INCORRECT TYPE. DISPOSE OF USED BATTERIES ACCORDING TO THE INSTRUCTIONS.

**ATTENTION:** RISQUE D'EXPLOSION SI LA BATTERIES EST REMPLAC'E PAR UN TYPE INCORRECT. REJETEZ LES BATTERIES UTILISE'ES SELON LES INSTRUCTIONS.

# TABLE OF CONTENTS

<b>SPECIFICATIONS</b> .....	<b>Page 1</b>
Standard Features .....	Page 1
Optional Features .....	Page 1
<b>SITE PREPARATION REQUIREMENTS</b> .....	<b>Page 2</b>
<b>INSTALLATION</b> .....	<b>Page 3</b>
Unpacking .....	Page 3
Mounting .....	Page 4
<b>INTERCONNECTIONS</b> .....	<b>Page 5</b>
Optional AC Power Adapter .....	Page 5
Height Rod Option – Detecto DHR .....	Page 5
Load Cell .....	Page 5
USB .....	Page 5
USB, Serial, Ethernet and WiFi Interface Specifications .....	Page 5
Serial I/O Port .....	Page 6
Ethernet (Optional) .....	Page 6
WiFi (Optional) .....	Page 6
<b>BATTERIES</b> .....	<b>Page 6</b>
Battery Status .....	Page 7
Low Battery Indicator .....	Page 7
Battery Charging .....	Page 7
Battery Installation/Replacement .....	Page 7
<b>KEYPAD FUNCTIONS</b> .....	<b>Page 8</b>
Basic Functions .....	Page 8
Body Mass Index Functions .....	Page 9
nControl Feature .....	Page 10
<b>ANNUNCIATORS</b> .....	<b>Page 11</b>
<b>OPERATION</b> .....	<b>Page 13</b>
Basic Operation .....	Page 13
Tare Storage and Recall .....	Page 15
Electronic Talley Roll (ETR) File .....	Page 16
Viewing the Electronic Talley Roll (ETR) File .....	Page 16
Basic Operation with ID .....	Page 17
Body Mass Index (BMI) Operation (Using Mechanical Height Rod or Manually Entering Height) .....	Page 19
Body Mass Index (BMI) Operation with ID (Using Mechanical Height Rod or Manually Entering Height) .....	Page 19
Body Mass Index (BMI) Operation with DHR (Using Digital Height Rod) .....	Page 20
Body Mass Index (BMI) Operation with ID and DHR (Using Digital Height Rod) .....	Page 21
<b>CALIBRATION SETUP (Protected Setup Parameters)</b> .....	<b>Page 22</b>
<b>CALIBRATION SETUP (Non-Protected Setup Parameters)</b> .....	<b>Page 33</b>
<b>SETUP REVIEW</b> .....	<b>Page 43</b>
<b>ERROR AND STATUS DISPLAYS</b> .....	<b>Page 45</b>
<b>BEFORE YOU CALL SERVICE</b> .....	<b>Page 45</b>
<b>CARE AND CLEANING</b> .....	<b>Page 46</b>
<b>SECURITY SEAL INSTALLATION</b> .....	<b>Page 46</b>
<b>PARTS IDENTIFICATION</b> .....	<b>Page 47</b>
<b>APPENDIX A (Top Plate Mounting Configurations)</b> .....	<b>Page 49</b>
<b>APPENDIX B (Connecting to a Welch Allyn CVSM Device)</b> .....	<b>Page 51</b>



# SPECIFICATIONS

Power Requirements . . . . .	6 "C" cell Ni-Cad or NiMH batteries ( <i>not included</i> ) OR an optional 12 VDC 1.25A wall plug-in AC power adapter (Cardinal part number MV2PWR).
Operating Temperature . . . . .	+5 to +40 °C
Display . . . . .	Weight: 5 digit, seven segment, 22.23 mm (7/8") high LCD Height: 4 digit, seven segment, 12.7 mm (1/2") high LCD BMI: 3 digit, seven segment, 12.7 mm (1/2") high LCD ID: 14 digit, 14 segment, 8.89 mm (.35") high LCD
Sensitivity . . . . .	0.83 uV/division (0 to 3.0 mV/V), Class III
Signal Input Range . . . . .	0 to 15mV max.4`
Transducer Excitation . . . . .	5.0 VDC
Number of Load Cells . . . . .	up to 4 each 350Ω
Load Cell Cable Length . . . . .	30 feet (9 meters) max.
Resolution . . . . .	6,000 divisions
Capacities . . . . .	6,000 divisions commercial
Division Value . . . . .	1, 2, 5 or 10 x 1, 0.1, 0.01, 0.001
Sample Rate . . . . .	1 to 100 samples per second selectable
Auto Zero Range . . . . .	0.5 or 1 through 9 divisions
Weighing Units . . . . .	Kilograms (kg)
Keyboard . . . . .	Membrane type with 19 keys
Enclosure Size . . . . .	248.9 mm W x 172.7 mm H x 45.7 mm D (9.8 inch W x 6.8 inch H x 1.8 inch D)
Construction . . . . .	Painted Steel

## Standard Features:

- Auto Shut-Off Feature
- Selectable Sleep-Mode
- StableSENSE<sup>®1</sup> Adjustable Filtering
- 99 Tare Storage
- Talley Roll
- Transaction Storage
- Battery Charging
- Serial I/O (Input/Output)
- USB Device Port
- Height Rod Port (for Optional Detecto Digital Height Rod)

## Optional Features:

- MVIP Ethernet Option
- MVWiFi Wireless Option
- Detecto DHR (Digital Height Rod)
- MV2PWR, 12VDC AC Power Adapter
- Welch Allyn<sup>®</sup> WACONNECT

<sup>1</sup> StableSENSE<sup>®</sup> is a digital filter utilizing proprietary software algorithms to remove or greatly reduce changes in the weight display resulting from movement on the scale platform. StableSENSE<sup>®</sup> can be used with clinical scales to lessen the effects of the patient's movement or vibration on the scale. Any application affected by vibration or movement on the scale platform can benefit using StableSENSE<sup>®</sup>.

# SITE PREPARATION REQUIREMENTS

The MedVue Model MV2 Medical Weight Analyzer is a precision weight indicating instrument. As with any precision instrument, it requires an acceptable environment to operate at peak performance and reliability. This section is provided to assist you in obtaining such an environment.


## Environmental

The MedVue Model MV2 Medical Weight Analyzer meets or exceeds all certification requirements within a temperature range of 14 to 104 °F (-10 to +40 °C).


In order to keep cooling requirements to a minimum, the MV2 should be placed out of direct sunlight and to provide adequate air circulation, keep the area around the MV2 clear.



Do not place the MV2 directly in front of a heating or cooling vent. Such a location will subject it to sudden temperature changes, which may result in unstable weight readings.



Insure that the MV2 has good, clean AC power and is properly grounded. In areas subject to lightning strikes, additional protection to minimize lightning damage, such as surge suppressors, should be installed.



## Electrical Power

The MedVue Model MV2 Medical Weight Analyzer has been designed to operate from a 100 to 240 VAC 50/60Hz 12 VDC 1.25A wall plug-in UL/CSA listed AC power adapter. Note that a special order is not required for operation at 230 VAC.

- The socket-outlet supplying power to the MV2 should be on a separate circuit from the distribution panel and dedicated to the exclusive use of the MV2.
- The socket-outlet shall be installed near the equipment and shall be easily accessible.
- The wiring should conform to national and local electrical codes and ordinances and should be approved by the local inspector to assure compliance.
- On installations requiring 230 VAC power, **it is the responsibility of the customer** to have a qualified electrician install the proper power adapter plug that conforms to national electrical codes and local codes and ordinances.



# INSTALLATION

## Unpacking

Carefully remove the MedVue Model MV2 Medical Weight Analyzer from the shipping carton and inspect it for any damage that may have taken place during shipment. You should keep the carton and packing material for return shipment if it should become necessary. The purchaser is responsible for filing all claims for any damages or loss incurred during transit.

Should your MedVue Model MV2 Medical Weight Analyzer come already installed on a scale, the following installation information does not apply to you.





Figure No. 1

## Mounting

The MedVue Model MV2 Medical Weight Analyzer is mounted to the scale (or an optional desktop/wall bracket) using a swivel ball and socket assembly.

To install the MV2 Rubber Ball in to the Swivel Ball Socket, slightly loosen the Adjustment Knob to allow the two parts of the socket to open enough to insert the Rubber Ball. Adjust the MV2 to the desired viewing angle and direction and then tighten the Adjustment Knob.



**NOTE:** The top plate mounting configuration varies depending on the model of your scale and may be different than the one shown in Figures No. 2 and 3. Refer to Appendix A for the various top plate mounting configurations.

## INTERCONNECTIONS

All input, output, power and option connections to the MV2 are made on the its rear panel. Connections for the Load Cell input and the Serial I/O are made using 9-pin "D" shaped connectors. The Height Rod option (Detecto DHR, Digital Height Rod) and Ethernet (MVIP) options use snap-in modular connectors. The USB port is a device (or upstream) port using an industry standard "B" connector. The optional 12VDC wall plug-in UL/CSA listed AC power adapter is connected using a power jack. Refer to Figure No. 1.

### Optional AC Power Adapter

To power the MV2 using the optional 12VDC wall plug-in AC power adapter, connect the plug from the adapter into the power jack on the back of the MV2 and then plug the power adapter into the proper electrical outlet. Refer to Figure No. 1. On models requiring 220 VAC, it is the customer's responsibility to obtain the correct power adapter plug.


### Height Rod – Detecto DHR (Digital Height Rod) Option

The optional Detecto DHR (Digital Height Rod) connects to the MV2 by a snap-in modular connector socket. Insert the modular connector of the DHR cable into the socket on the rear of the MV2 it locks in place (a clicking sound will be heard when it is locked in place).

### Load Cell

The load cell cable connects to the MV2 via a 9-pin "D" connector on the rear panel of the MV2. Figure No. 4 shows the pin identification for the load cell connector. Make certain that the pins are correctly identified before soldering a wire to them. Use the connector retaining screws to hold the load cell cable connector securely to the rear panel.

<u>PIN NO.</u>	<u>FUNCTION</u>
1	+EXCITATION
2	-SIGNAL
5	SHIELD
6	-EXCITATION
7	+SIGNAL



Note that pins 3, 4, 8 and 9 are not used.

Figure No. 4

### MATING CONNECTOR INFORMATION

<b>DESCRIPTION</b>	<b>ITEM</b>	<b>Cardinal Part #</b>
CONNECTOR	DE9-P	6610-2379
CONNECTOR SHELL	C883010001	6610-1131

### USB

The USB port on the MV2 is a device (or upstream) port and uses readily available cables with the industry standard "B" connector.

### USB, Serial, Ethernet and WiFi Interface Specifications

The USB, Serial and optional Ethernet and WiFi interfaces on the MV2 can be configured during the setup and calibration procedure or during the setup review operation. Using either method, it is possible to select the operation of the interfaces.

The interfaces may be connected to a computer for transmission of weight and associated data to a PC-based EMR (electronic medical record) software program. The data can be transmitted on demand (pressing the **PRINT/ENTER** key) or on receipt of a command from the computer.

## Serial I/O Port

The serial I/O port is a 9-pin "D" connector on the rear panel of the MV2. Figure No. 5 shows the Serial I/O connector along with the identity of the pins used. Make certain that the pins are correctly identified before soldering a wire to them. Use the connector retaining screws to hold the serial cable connector securely to the rear panel.

<u>PIN NO.</u>	<u>FUNCTION</u>
2	DATA INPUT (RXD)
3	DATA OUTPUT (TXD)
5	SIGNAL GROUND (GND)
9	VCC or 5VDC




Figure No. 5

Note that pins 1, 4 and 6 through 8 are not used.

The MV2 serial interface can be configured during the setup and calibration procedure or during the setup review operation. Using either method, it is possible to select the operation of the serial interface as well as select the baud rate.

- The baud rates supported are: 2400, 4800, 9600, 14.4k, 19.2K, 28.8K, 38.4K, 56K, 57.6K, and 115.2K baud.
- The data format is fixed at 8 bits, No parity, and 1 stop bit.
- The MV2 is shipped from the factory with the baud rate set to 9600 baud.

## Ethernet (Optional)

The MVIP Ethernet option requires the purchase of the optional Ethernet Interface Option. Once installed, connections to the MV2 are by CAT 5 or equivalent cable. Consult your network administrator for the proper procedure to terminate and connect the cable.

## WiFi (Optional)

The MVWiFi option requires the purchase of the optional WiFi Interface Option. Once installed, connections to the MV2 are by the wireless network. Consult your network administrator for instructions to connect the MV2 to your wireless network.

## BATTERIES

The MV2 can use 6 "C" size Ni-Cad or NiMH batteries (*not included*). You must first obtain and install batteries before operations can begin. Batteries are contained in a battery holder inside the MV2. Access is via a removable panel on the back of the MV2. Refer to Figure No. 1.




**CAUTION!** The MV2 has internal circuitry that when used in conjunction with the AC power adapter, recharges the Ni-Cad or NiMH batteries. Because the MV2 has this charging capability, you **CAN NOT** use non-rechargeable batteries in it.



**NOTE:** All 6 batteries must be of the same type. They must be all Ni-Cad or all NiMH. **DO NOT** mix Ni-Cad or NiMH batteries.

## Battery Status Symbol



A battery status symbol will be displayed when batteries are installed. The number of bars indicates the level of charge remaining in the batteries.

4 Bars = 100%
3 Bars = 75%
2 Bars = 50%
1 Bar = 25%
No Bars = Low Battery

## Low Battery Indicator

When the batteries are near the point they need to be recharged, the battery status symbol will not show any bars. If the battery voltage drops too low for accurate weighing, the MV2 will automatically shut off and you will be unable to turn it back on.

## Battery Charging

To recharge the Ni-Cad or NiMH batteries, the AC power adapter must be connected to an AC power outlet and plugged into the indicator. It will take approximately 6 hours to fully recharge the batteries in the indicator. While the batteries are charging the indicator can still be operated. Note that charging the batteries for more than 6 hours *will not* damage them.

## Battery Installation/Replacement

To install or remove the batteries, the following steps should be followed:

1. Make sure the AC power adapter is unplugged.
2. Remove the MV2 from the mounting bracket and turn it over so that the display is facing away from you.
3. Locate the rectangular panel on the back of the MV2 and remove the thumb screw. Refer to Figure No. 1.
4. Remove the panel (lift straight up and slide it out) exposing the battery holder.
5. If installing new batteries, proceed to step 6. If replacing the batteries, remove all 6 batteries from the battery holder and then proceed to step 6.
6. Install the new 6 "C" size batteries in the holder, noting the polarity markings located in the battery holder. Refer to Figure No. 6.
7. After placing all 6 batteries in holder, replace panel on back of the MV2 (slide tab into slot on rear panel) and install thumb screw.
8. Turn the MV2 over (display facing up) and press the **ON/OFF** key.
9. If display turns on, batteries have been installed correctly. If not, remove panel and check for one or more improperly positioned batteries.
10. Return the MV2 to the mounting bracket.
11. The MV2 is now ready for operation.





Figure No. 6

# KEYPAD FUNCTIONS



**DO NOT** operate the keypad with pointed objects (pencils, pens, etc).  
 Damage to keypad resulting from this practice is **NOT** covered under warranty.




**Figure No. 7**



This is the **ON/OFF** key. With the MV2 off, pressing this key will apply power to it and turn on the display. If the MV2 is already on, pressing the key will turn it off.

## Basic Functions



This is the **ZERO** key. It is used to reset the display to zero up to the selected limit of either 4% or 100% of the scale capacity. The zero limit is set during setup and calibration of the MV2.



This is the **UNITS** key. It is not currently used.


## Basic Functions, Cont.




This is the **LOCK/RELEASE** key. It is used to lock and unlock the display. If the HOLD feature was enabled during setup and calibration, pressing this key (after obtaining a stable weight) will cause the MV2 to lock onto the weight. Pressing this key again will unlock the display and return it to zero.

If the AUTO LOCK feature was enabled, the weight display will lock after placing a load on the scale and obtaining a stable weight value. Pressing this key will unlock the display and return the weight display to zero, or if another load is applied to the scale, lock onto the new weight after obtaining a stable weight value. **NOTE:** The lock feature is for non-commercial (NOT "Legal for Trade") applications.

## Body Mass Index (BMI) Functions



This is the **NET/GROSS** key. It is used to toggle between Net and Gross weight modes. The selected mode is indicated by turning on the appropriate annunciator on the display. Note that if no valid tare weight has been entered; pressing this key will cause (No Tare) to be shown in the 14-segment display.



This is the **TARE** key. It is a dual function key. Pressing it, selecting a memory location and pressing the **PRINT/ENTER** key (Pushbutton Tare mode) will cause the current gross weight to be stored as the new tare weight. Pressing this key, entering a numeric weight value (Keypad Tare), selecting a memory location and then pressing the **PRINT/ENTER** key will cause the value entered to be stored as the new tare weight. Either method will change the weight display to the net weight display mode and turn on the Net annunciator.

### Pushbutton Tare Function:


Press the **TARE** key and then the **ID/HEIGHT** key, input a number between 0 and 99 and then press the **PRINT/ENTER** key to store current Gross weight in a memory location as a recallable tare weight.

### Keypad Tare Function:

Press the **TARE** key, input the weight value, press the **ID/HEIGHT** key, input a number between 0 and 99 and then press the **PRINT/ENTER** key to store the entered value in a memory location as a recallable tare weight.



**NOTE:** Tare weights equal to or greater than scale capacity cannot be entered. In addition, the keypad tare weight division value must be the same as the scale division value. For example, a unit with .05 kg as the division value will display **WRONG GRAD5** if you attempt to enter 1.03 for the tare weight.




This is the **ID/HEIGHT** key. This key is used to display and enter an identification (ID) use in the ID Storage feature and to enter the height (feet and inches or centimeters) for the Body Mass Index (BMI) calculation.


Pressing this key will change the display to show the currently selected identification (ID) and allow the entry of a new ID. The identification (ID) can be up to 14 alphanumeric characters.

To recall the stored tare weight from the memory location, press the **ID/HEIGHT** key, input a number between 1 and 99, and press the **TARE** key.

## Body Mass Index (BMI) Functions



This is the **CLEAR (BACKSPACE)** key. It is used to clear an incorrect entry from the display without processing the data. If an incorrect entry is made, press this key and re-enter the correct data. When entering the ID, pressing this key will act as a “backspace” key (it will back up one space and delete the character in that space). **NOTE:** The **CLEAR** key must be pressed before the **PRINT/ENTER** key to ensure the data is not processed.



This is the **PRINT/ENTER** key. It serves several purposes. First, it is used to signal completion of the data entry and causes the MV2 to process the data entered. Second, it is used to initiate the transmission of weight and associated data using the selected I/O port.



**NOTE:** The MV2 will not respond to pressing the **PRINT/ENTER** key unless the weight display is stable. If displaying gross weight, only the gross weight is printed. If displaying net weight, the gross, tare, and net weights are printed. If the ID and BMI features are used, their data is printed along with the weight data.



These keys are used to enter alphanumeric data during the setup and calibration as well as during normal operation of the MV2.

**NOTE:** The **0** and **1** keys have dual functions. They are used to enter numeric data during setup and calibration as well as during normal operations. In addition, they are used to select YES or NO to various prompts.



To enter letters, you must press the key in succession until the desired letter is displayed. This is similar to the way a cell phone keypad operates. (Ex: If you want to enter the letter C, you would need to press the **2** key four times).

## nControl Feature

The MV2 includes support for *nControl*. *nControl* is a PC based program that can design a ticket then download the ticket information to the MV2. The MV2 allows 1 programmable format (up to 2kb in size) instead of the ASCII format default ticket.

*For more information on nControl, refer to the nControl Fast Start Guide.*

Time, dd/mm/yy

ID: xxxxxxxxxxxx  
 BMI: xx.x  
 Height: xx.x cm (or mm)  
 Weight: xxx.xuu G  
 xxx.xuu PT  
 xxx.xuu N

**SAMPLE TICKET**



**NOTE:** If the ID/HEIGHT key is disabled, the appropriate field will be left out when printing a ticket.



# ANNUNCIATORS

The annunciators are turned on to indicate that the display is in the mode corresponding to the annunciator label or that the status indicated by the label is active.




Figure No. 8

## GROSS

The GROSS annunciator is turned on to show that the weight displayed is the gross weight. Gross weight will be displayed when no tare weight value is stored.

## TARE

The TARE annunciator is turned on to show that the MV2 is in a weight mode in which a known tare (container) weight value is stored.

## NET

The NET annunciator is turned on to show that the weight displayed is the net weight. Net weight is determined by subtracting the stored tare weight from the gross weight. The tare weight is usually the weight of the empty container. Note that the NET annunciator is only active when a tare weight value is stored.

## ▲▲ (STABLE)

The (▲▲) annunciator is turned on when the weight display is stable. When off, it means that the change in successive weight samples is greater than the motion limits selected during setup and calibration of the MV2.

## →0← (ZERO)

The (→0←) annunciator is turned on to indicate that the weight is within +/- 1/4 division of the center of zero.

## 🔒 (LOCK)

The (🔒) annunciator is turned on to show that the MV2 is locked onto the displayed weight. In operation after obtaining a stable weight value, pressing the **LOCK/RELEASE** key (or resetting the Auto Lock) will cause the MV2 to lock onto the weight and turn on the annunciator. Pressing the **LOCK/RELEASE** key a second time (or dropping below the Auto Lock reset value) will unlock the display and turn off the annunciator. **NOTE:** The lock feature (Hold Mode) must be enabled during Calibration Setup.

## **CAL**

The CAL annunciator will be turned on after pressing the calibration switch, to show that the MV2 is in the Setup and Calibration mode.

## **kg**

The kg annunciator is turned on to indicate that the displayed weight is in kilograms.

## **cm**

The cm annunciator is turned on to indicate that the displayed height measurement is in centimeters.

## **mm**

The mm annunciator is turned on to indicate that the displayed height measurement is in millimeters.

## **AM**

The AM annunciator is turned on to indicate that the displayed time is before noon.

## **PM**

The PM annunciator is turned on to indicate that the displayed time is after noon.




The 🕒 annunciator is turned on to indicate that the displayed value is the time.

## **BMI (Body Mass Index)**


The BMI annunciator is turned on when displaying the calculated body fat.




**WiFi**



**RS232**



**Ethernet**



**USB**

These annunciators are turned on when optional hardware is installed and in use.

# OPERATION

## Basic Operation

### To Weigh

1. Press **ON/OFF** key to turn the MV2 on.
2. Press **ZERO** key to zero weight display. ZERO, GROSS and kg annunciator will turn on to show that the scale is ready for use.
3. Place patient on scale and read weight display.
4. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
5. Remove patient from scale.

### Zero Weight Display

1. In Gross Weight mode (GROSS annunciator on), press **ZERO** key.
2. Weight display will return to zero. ZERO and STABLE ▲▲ annunciators will turn on to show a stable, center-of-zero (→0←) gross weight condition.

## Basic Tare Operation

### Push Button Tare (Tare weight is NOT known)

**IMPORTANT! Tare entry is not allowed if Auto Hold is enabled.**

1. In Gross Weight mode (GROSS annunciator on), place item (wheelchair, walker, etc...) to be tared on scale.
2. Press the **TARE** key.
3. The Display will change to show ENTER TARE WT, the weight of item on scale and the TARE annunciator will turn on.
4. Press **PRINT/ENTER** key.
5. Display will change to 0.0 and NET annunciator will turn on to show that net weight is being displayed. The item's weight has been entered as "tare weight".
6. Place patient on scale
7. Read weight display.
8. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
9. Remove patient from scale.

### To Clear the Tare Weight

1. Remove all material from scale platform and press **TARE** key.
2. Press **ZERO** key.
3. This will clear the tare weight and return the MV2 to the Gross weight mode.

### **Tare Weight Entry (Tare weight is known)**

1. In Gross Weight mode (GROSS annunciator on), press the **TARE** key.
2. The Display will change to show ENTER TARE WT and the TARE annunciator will turn on.
3. Use the numeric keys and enter a tare weight. Note that when entering tare weights, a maximum of 4 digits can be entered and the numbers advance from right to left in the display. The number of leading or trailing zeros required to obtain your desired tare is dependent upon the DIVISION VALUE selected in the setup procedure.

For example:	Division Value:	0.2 kg
	Desired Tare Value:	10.6 kg
	Key Sequence:	1 0 6

4. After the tare weight has been entered, press **PRINT/ENTER** key.
5. The display will show a minus Net weight and NET annunciator will turn on.
6. Place patient on scale
7. Read weight display.
8. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
9. Remove patient from scale.

### **Tare Weight Recall**

1. In Net Weight mode (NET annunciator on), press **TARE** key.
2. The Display will change to show ENTER TARE WT and the previously entered tare weight. Note that the TARE annunciator will turn on.
3. Press **PRINT/ENTER** key to return to Net Weight mode (NET annunciator on).

### **To Clear the Tare Weight**

1. Remove all material from scale platform and press **TARE** key.
2. Press **ZERO** key.
3. This will clear the tare weight and return the MV2 to the Gross weight mode.

## Tare Storage and Recall

**NOTE: ID must be enabled to use stored Tare. Refer to SYSTEM Menu, Enable ID.**

### To Store Tare Weight

1. In Gross Weight mode (GROSS annunciator on), press the **TARE** key.
2. The Display will change to show ENTER TARE WT and the TARE annunciator will turn on.
3. Use the numeric keys and enter a tare weight. Note that when entering tare weights, a maximum of 4 digits can be entered and the numbers advance from right to left in the display. The number of leading or trailing zeros required to obtain your desired tare is dependent upon the DIVISION VALUE selected in the setup procedure.

For example:	Division Value:	0.2 kg
	Desired Tare Value:	10.6 kg
	Key Sequence:	1 0 6

4. After the tare weight has been entered, press **ID/HEIGHT** key.
5. Display will change to show *id*.
6. Using numeric keys enter the memory location (1 to 99).
7. Press **PRINT/ENTER** key.
8. The tare weight is now stored in memory and the display will show a minus Net weight and NET annunciator will turn on.
9. Place patient on scale
10. Read weight display.
11. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
12. Remove patient from scale.

### To Recall Tare Weight

1. Press **ID/HEIGHT** key.
2. Display will change to show *id*.
3. Using numeric keys enter the memory location (1 to 99).
4. Press **TARE** key.
5. The display will change to show ENTER TARE WT and the previously entered tare weight. Note that the TARE annunciator will turn on.
6. If tare weight displayed is acceptable, press **PRINT/ENTER** key to save it.
7. Otherwise, press **ZERO** key to clear the tare weight and return the MV2 to the Gross weight mode.

## Electronic Talley Roll (ETR) File

The MV2 is equipped with an Electronic Talley Roll (ETR) feature that stores the weight and a consecutive transaction number each time the indicator receives a Talley command from the computing peripheral. The data is maintained in a customer supplied Micro SD card is available for use by a weights and measures inspector.



**NOTE:** The Micro SD card is inserted into the battery charger board under the right end cap (as viewed from the MV2 back). See Figure No. 9.



**CAUTION!** Do not plug or unplug the battery charger board while the MV2 has power (batteries installed or AC power adapter plugged in). **DAMAGE TO THE MICRO SD CARD WILL OCCUR.**

When enabled, each time the indicator receives an ETR request (SMA XT command), the gross weight and an internal consecutive transaction number (that cannot be modified) will be stored in memory. The ETR file is structured such that when the capacity of the Micro SD card is reached, the newest transaction will replace the oldest. The ETR file can only be displayed and cannot be manually cleared or printed.




Figure No. 9

## Viewing the Electronic Talley Roll (ETR) File

1. With the indicator in normal operations mode, press the **ID/HEIGHT** key.
2. Display will change to show *id*.
3. Using the numeric keys, enter the Talley Roll number.
4. Press the **UNITS** key.
5. The display will return to the Gross weight mode and show the weight associated with the Talley Roll number in the Message Display at the bottom of the MV2 display.  
**NOTE:** If the Talley Roll number is not found or the Micro SD card is not installed, the weight display will show *ERR* for one second and return to the ID (*id*) prompt.
6. After displaying the record for 3 seconds, the display will return to normal operation.

## Basic Operation with ID

### To Weigh

1. Press **ON/OFF** key to turn the MV2 on.
2. Press **ZERO** key to zero weight display. ZERO, GROSS and kg annunciator will turn on to show scale is ready for use.
3. Press **ID/HEIGHT** key.
4. Display will change to show *Id*.
5. Using numeric keys enter up to an 14 digit alphanumeric identification number.
6. Press **PRINT/ENTER** key.
7. Place patient on scale
8. Read weight display.
9. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
10. Remove patient from scale.

### Push Button Tare (Tare weight is NOT known)

**IMPORTANT! Tare entry is not allowed if Auto Hold is enabled.**

1. In Gross Weight mode (GROSS annunciator on), place item (wheelchair, walker, etc.) to be tared on scale.
2. Press **TARE** key.
3. Display will show weight of item on scale and the TARE annunciator will turn on.
4. Press **PRINT/ENTER** key.
5. Display will change to zero and NET annunciator will turn on to show that net weight is being displayed. The item's weight has been entered as "tare weight".
6. Press **ID/HEIGHT** key.
7. Display will change to show *Id*.
8. Using numeric keys enter up to an 14 digit alphanumeric identification number.
9. Press **PRINT/ENTER** key.
10. Place patient on scale
11. Read weight display.
12. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
13. Remove patient from scale.

### **Tare Weight Entry (Tare weight is known)**

1. In Gross Weight mode (GROSS annunciator on), press **TARE** key.
2. Display will show *0.0* or previously entered tare and TARE annunciator will turn on.
3. If weight displayed is acceptable, press **PRINT/ENTER** key to save it. Otherwise, use the numeric keys and enter a new tare weight. Note that when entering tare weights, a maximum of 4 digits can be entered and the numbers advance from right to left in the display. The number of leading or trailing zeros required to obtain your desired tare is dependent upon the DIVISION VALUE selected in the setup procedure.

For example:            Division Value:            0.2 kg  
                                  Desired Tare Value:        10.6 kg  
                                  Key Sequence:             1 0 6

4. After the new tare weight has been entered, press **PRINT/ENTER** key.
5. Display will show a minus Net weight and NET annunciator will turn on.
6. Press **ID/HEIGHT** key.
7. Display will change to show *id*.
8. Using numeric keys enter up to an 14 digit alpha numeric identification number.
9. Press **PRINT/ENTER** key.
10. Place patient on scale
11. Read weight display.
12. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
13. Remove patient from scale.



## Body Mass Index (BMI) Operation

### Using Mechanical Height Rod or Manually Entering Height

1. Press **ON/OFF** key to turn the MV2 on.
2. Press **ZERO** key to zero weight display. ZERO, GROSS and lb or kg annunciator will turn on to show scale is ready for use.
3. Perform tare operation if required. Refer to Push Button Tare or Tare Weight Entry in previous section for instructions on using tare.
4. Press **ID/HEIGHT** key.
5. If Centimeters was selected for HEIGHT UNITS in setup, the display will change to show CENTIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # . #) for height in centimeters and then press **PRINT/ENTER** key.
6. If Millimeters was selected for HEIGHT UNITS in setup, display will change to show MILLIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # #) for height in millimeters and then press **PRINT/ENTER** key.
7. Place patient on the scale.
8. Read weight and BMI displays.
9. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket. Note that BMI and height display will clear when ticket prints.
10. Remove patient from scale.
11. If a printer is not connected to scale, BMI display will clear when patient is removed from scale and weight returns to zero.

## Body Mass Index (BMI) Operation with ID

### Using Mechanical Height Rod or Manually Entering Height

1. Press **ON/OFF** key to turn MV2 on.
2. Press **ZERO** key to zero weight display. ZERO, GROSS and lb or kg annunciator will turn on to show scale is ready for use.
3. Press **ID/HEIGHT** key.
4. Display will change to show *Id*.
5. Using numeric keys enter up to a 14 digit alphanumeric identification number.
6. Press **PRINT/ENTER** key.
7. If Centimeters was selected for HEIGHT UNITS in setup, the display will change to show CENTIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # . #) for height in centimeters and then press **PRINT/ENTER** key.
8. If Millimeters was selected for HEIGHT UNITS in setup, display will change to show MILLIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # #) for height in millimeters and then press **PRINT/ENTER** key.
9. Place patient on the scale.
10. Read weight and BMI displays.
11. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket. Note that BMI, ID and height display will clear when ticket prints.
12. Remove patient from scale.
13. If a printer is not connected to scale, BMI display will clear when patient is removed from scale and weight returns to zero.

## Body Mass Index (BMI) Operation with DHR

### Using Digital Height Rod



**IMPORTANT!** Before performing the operations on this page, the digital height rod must be in its “starting position” (inner sliding tube must be down completely inside outer stationary tube and headpiece folded flat against stationary tube) prior to turning on the MV2.

1. Make sure height rod is in starting position.
2. Press **ON/OFF** key to turn MV2 on.
3. Press **ZERO** key to zero weight display. ZERO, GROSS and kg annunciator will turn on to show scale is ready for use. Note that height reading will show 110.0 cm or 1100 mm (depending on the HEIGHT UNITS setting in Setup).
4. Perform tare operation if required. Refer to Push Button Tare or Tare Weight Entry in previous section for instructions on using tare.



**NOTE:** If patient is shorter than 110 cm (1100 mm) or taller than 201.2 cm (2012 mm), proceed to step 5 or 6 depending on the HEIGHT UNITS setting. Otherwise, press the **PRINT/ENTER** key without entering a value for steps 5 or 6 and proceed to Step 7.

5. Press **ID/HEIGHT** key. If Centimeters was selected for HEIGHT UNITS in setup, the display will change to show CENTIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # . #) for height in centimeters and then press **PRINT/ENTER** key.
6. Press **ID/HEIGHT** key. If Millimeters was selected for HEIGHT UNITS in setup, display will change to show MILLIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # #) for height in millimeters and then press **PRINT/ENTER** key.
7. Place patient on the scale.
8. Grasp height rod at hinge pin (see Figure No. 10) and raise it well above patient’s head and then lift headpiece to horizontal position. Carefully lower height rod until headpiece rests on top of patient’s head.
9. BMI will automatically be displayed with height.
10. Read weight, height and BMI displays.
11. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
12. While holding headpiece horizontal, raise height rod well above patient’s head.
13. Remove patient from scale and return height rod to starting position.
14. If Tare operation was performed, press **NET/GROSS** key to return to Gross weight mode.
15. Press **ZERO** key. This will reset tare weight to zero.
16. BMI display will clear when weight returns to zero.



**Figure No. 10**

## Body Mass Index (BMI) Operation with ID and DHR

### Using Digital Height Rod



**IMPORTANT!** Before performing the operations on this page, the digital height rod must be in its “starting position” (inner sliding tube must be down completely inside outer stationary tube and headpiece folded flat against stationary tube) prior to turning on the MV2.

1. Make sure height rod is in starting position.
2. Press **ON/OFF** key to turn MV2 on.
3. Press **ZERO** key to zero weight display. ZERO, GROSS and kg annunciator will turn on to show scale is ready for use. Note that height reading will show 100.0 cm or 1000 mm (depending on the HEIGHT UNITS setting in Setup).
4. Perform tare operation if required. Refer to Push Button Tare or Tare Weight Entry in previous section for instructions on using tare.
5. Press **ID/HEIGHT** key.
6. Display will change to show **Id=**.
7. Using numeric keys enter up to a 14 digit alpha numeric identification number.
8. Press **PRINT/ENTER** key.



**NOTE:** If patient is shorter than 3' 7.3" (110 cm) or taller than 6' 7.3" (201.2 cm), proceed to step 9, 10 or 11 depending on the HEIGHT UNITS setting. Otherwise, press the **PRINT/ENTER** key without entering a value for steps 9 or 10 and proceed to Step 11.

9. If Centimeters was selected for HEIGHT UNITS in setup, the display will change to show CENTIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # . #) for height in centimeters and then press **PRINT/ENTER** key.
10. If Millimeters was selected for HEIGHT UNITS in setup, display will change to show MILLIMETERS.
  - a. Use numeric keys to enter up to 4 digits (# # # #) for height in millimeters and then press **PRINT/ENTER** key.
11. Place patient on scale.
12. Grasp height rod at hinge pin (see Figure No. 11) and raise it well above patient's head and then lift headpiece to horizontal position. Carefully lower height rod until headpiece rests on top of patient's head.
13. BMI will automatically be displayed with height.
14. Read weight, height and BMI displays.
15. If a printer is connected to scale, press **PRINT/ENTER** key to print a ticket.
16. While holding headpiece horizontal, raise height rod well above patient's head.
17. Remove patient from scale and return height rod to starting position.
18. If Tare operation was performed, press **NET/GROSS** key to return to Gross weight mode.
19. Press **ZERO** key. This will reset tare weight to zero.
20. BMI display and ID will clear when weight returns to zero.



**Figure No. 11**


# CALIBRATION SETUP

## (Protected Setup Parameters)

Your MedVue Model MV2 Medical Weight Analyzer has been thoroughly tested and calibrated before being shipped to you. If you received it with a scale, calibration is not necessary. If the MV2 is being connected to a scale for the first time or recalibration is necessary for other reasons, the following describes the steps to perform calibration setup on the MV2.

### To Enter Calibration Setup:

1. With the power off, remove the two (2) screws from the right end cap (as viewed from the MV2 back) and remove the end cap.
2. Locate the calibration switch (access is through a hole near the bottom of the side panel. See Figure No. 12).
3. Press the **ON/OFF** key to turn the MV2 on.
4. The MV2 will scroll the model name and the software version and then change to the Gross weight mode.
5. With the MV2 on, press and hold the calibration switch until the display changes to show:



Calibration Switch

Figure No. 12

6. Release the calibration switch.
7. The MV2 is now ready for setup and calibration

During setup and calibration it is necessary to enter operational parameters via the MV2's keypad. The following keys and the numeric keypad are used to input the values.




**NOTE: The membrane keypad is not to be operated with pointed objects (pencils, pens, fingernails, etc). Damage to keypad resulting from this practice will NOT be covered under warranty.**

### Navigation Keys

The Navigation Keys are used to move around the menu selections during setup and calibration. They are used to select the parameter to change and to advance to the next parameter, backup to the previous parameter or "exit" to the previous screen.

### GROSS/NET – (Previous Menu or Parameter)



If at the menu level, this will take you to the next menu item in the list below. If **TIME SETTINGS** is currently displayed and the setup has been accessed through the calibration switch, then this will return you to **SCALE**. If setup has been accessed through setup review, then pressing this when **TIME SETTINGS** is displayed will return you to **SYSTEM**. If at the parameter level, the exact item this takes you to may depend on a [parameter condition].

## Navigation Keys, Cont.

### ID/HEIGHT – (Next Menu or Parameter)




If at the menu level, this will take you to the previous menu item in the list below. If SCALE is currently displayed and the setup has been accessed through the calibration switch, then this will return you to TIME SETTINGS. If setup has been accessed through setup review, then pressing this when SYSTEM is displayed will return you to TIME SETTINGS. If at the parameter level the exact item this takes you to may depend on a [parameter condition].

### UNITS – (Go up a level)



If at the parameter level, this returns you to the menu level. If at the menu level, this returns you to normal operation

### PRINT/ENTER – (Go into Menu or save Parameter)



If at the menu level, this will take you to the first parameter in the currently displayed menu in the list below. If at the parameter level, this will store the currently displayed value for that parameter and proceed to the next parameter in the list below, which may depend on a [parameter condition].

## Numeric Keys

The Numeric Keys are used to input values for the Calibration Setup parameters. Some parameter prompts require only a yes or no selection, while others require a numeric value to be entered.

**NOTE:** The 1 and 0 keys have dual functions. They are used to enter numeric values as well as to select YES or NO to various parameter prompts.



## Calibration Setup Menu

SCALE	Scale Configuration. Proceed to the Scale Capacity parameter.
ZERO	Configure Zero functions. Proceed to the Zero Limit parameter.
FILTER	Analog to Digital Filtering. Proceed to the Filter Mode parameter.
CALIBRATION	Perform Calibration. Proceed to the Calibrate parameter.
GRAVITY	Gravity Constant Function. Proceed to the Enable Gravity parameter.
LOCK	Lock/Release feature. Proceed to the Hold Mode parameter.
SYSTEM	Operational Selections. Proceed to the Language parameter.
SERIAL PORT	Serial I/O Configuration. Proceed to the RS232 Mode parameter.
USB PORT	USB Port Configuration. Proceed to the USB Mode parameter.
ETHERNET/WIFI	Ethernet/WiFi Configuration. (Displayed if option card is present). Proceed to the Ethernet Mode parameter.
TIME SETTINGS	Time and Date Configuration. Proceed to the Year parameter.

## SCALE

With **SCALE** displayed, press the **PRINT/ENTER** key. The display will change to **CAPACITY**. Proceed to the **CAPACITY** (Scale Capacity) parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the **ZERO** menu.

## CAPACITY (Scale Capacity)

The display will show **CAPACITY** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 through 65,535.

## OIML (OIML Compliance)

This setting determines whether the MV2 will be used in an application requiring an OIML Compliance.

The display will show **OIML** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = OIML Compliance is required      no = OIML Compliance is NOT required

Selecting YES enables viewing data from the Electronic Talley Roll (ETR) file using the MV2 display. Refer to page 16. Note that a customer supplied Micro SD card must be installed in the battery charger board to perform this function. In addition, the **OVER CAPACITY** setting is set to **CAP + 9** grads

## INTERVAL (Interval Setting)

The display will show **INTERVAL** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1, 2 or 5.

## dECIMAL (Decimal Point Precision)

The display will show **dECIMAL** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

0 = #####      1 = #####.#      2 = #####.##      3 = ###.###

## MOTION RANGE (Motion Range)

The motion range is the number of divisions of change permitted before indicating unstable (the **STABLE** annunciator turns off).

The display will show **MOTION RANGE** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 9.

### **SAMPLE RATE (Sample Rate)**

The display will show **SAMPLE RATE** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

The sample rate may be set from a minimum of 1 sample per second to a maximum of 100 samples per second in one sample per second intervals.

Allowable values are 1 through 100.

## ZERO

With ZERO displayed, press the **PRINT/ENTER** key. The display will change to ZERO LIMIT. Proceed to the Two Percent Zero Tracking Range Limit parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the FILTER menu.

### ZERO LIMIT (Two Percent Zero Tracking Range Limit)

This setting limits the scale to zero only + or - 2% of the scale capacity.

If enabled, (ZERO LIMIT = YES) zero tracking and push button zero is limited to 2% of scale capacity.

If disabled, (ZERO LIMIT = no) zero tracking and push button zero will operate up to full scale capacity.

The display will show ZERO LIMIT and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = 2% Zero Limit is Enabled

no = 2% Zero Limit is Disabled



**NOTE: If the MV2 is used in an Legal for Trade application, the Zero Limit must be set to YES (ZERO LIMIT = YES).**

### POWER UP ZERO (Power Up Zero)

The display will show POWER UP ZERO and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = Weight display will be reset to zero automatically on power up

no = Weight display will not be reset to zero on power up

### ZERO TRACKING (Zero Tracking)

The zero tracking is a value in scale divisions that will automatically zero the scale within the limits determined by the setting for this prompt.

The display will show ZERO TRACKING and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 18 (1 to 9 divisions by 0.5 divisions) e.g. 1 = .5 grads, 2 = 1 grad, 3 = 1.5 grads, etc...

Note that entering a 0 (zero) will disable zero tracking.



## FILTER

With FILTER displayed, press the **PRINT/ENTER** key. The display will change to FILTER MODE. Proceed to the Filter Type parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the CALIBRATION menu.

### FILTER MODE (Filter Type)

The display will show FILTER MODE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1 or 2.

- 0 = None Disabled, No Filtering
- 1 = IIR Infinite Impulse Response – (slower response but better for higher capacity scales)
- 2 = FIR Finite Impulse Response – (faster response, better for filtering out vibration)

**NOTE:** The following prompt, FILTER LEVEL will only be displayed if you select option 1 or 2 for FILTER MODE.

### FILTER LEVEL (Filter Level Selection – Only if FILTER MODE = 1 or 2)

The display will show FILTER LEVEL and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1 or 2.

- 0 = Minimal Filtering
- 1 = Moderate Filtering
- 2 = Custom Filtering

**NOTE:** The following prompts, FILTER bREAK (Filter Break Range) and FILTER VALUE (Filter Value) will only be displayed if you select 2 (Custom Filtering) for FILTER LEVEL.

### FILTER bREAK (Filter Break Range – Only if FILTER LEVEL = 2)

The filter break range is a number from 1 to 64 that corresponds to the number of division changes to break out of filtering.

The display will show FILTER bREAK and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 through 64.

### FILTER VALUE (Filter Value – Only if FILTER LEVEL = 2)

The filter level is a number from 1 to 16 that corresponds to the level of filtering with 1 being the least and 16 being the greatest filtering.

The display will show FILTER VALUE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 through 16.

## CALIBRATION

With CALIBRATION displayed, press the **PRINT/ENTER** key. The display will change to CALIBRATE. Proceed to the Perform Calibration parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the GRAVITY menu.

### CALIBRATE (Perform Calibration)

The display will show CALIBRATE and the current setting (NO).

- If the scale has been previously calibrated and you wish to skip calibration (and retain the internal calibration factor), press the **PRINT/ENTER** key. The display will change to GRAVITY ENABLE. Proceed to the GRAVITY menu, Gravity Enable parameter.
- If calibration is desired, press the **1/YES** key to select YES and then press the **PRINT/ENTER** key. The display will change to NUM CAL POINTS.

Allowable values are YES or no.

YES = Perform Calibration and  
proceed to NUM CAL POINTS

no = Skip Calibration and proceed to  
the GRAVITY ENABLE parameter

### NUM CAL POINTS (Number Of Calibration Points – Only if CALIBRATE = YES)

The MedVue Model MV2 Medical Weight Analyzer has two (2) modes that can be used to perform calibration. The calibration modes are as follows:

#### Single Calibration Point

This is a standard calibration method requiring one weight, an empty scale and has one conversion factor. This method uses two calibration points (a test load on the scale and an empty scale) to establish a zero (no load) calibration value and to span the MV2.

#### Multi-point Calibration

This method requires two weights, an empty scale and has two conversion factors. It uses three calibration points, (a load on the scale, half the previous load and an empty scale). The three points correspond to zero weight, mid-point weight and test load (weight).

The display will show NUM CAL POINTS and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 or 2.

1 = Single Calibration Point

2 = Multi-Point Calibration

# Single Calibration Point

## LOADd CAL WT (Load Calibration Weight)

The display will show LOADd CAL WT (which is a prompt for entry of the calibration weight value) and placement of this amount of test weight on scale platform.

1. Make certain scale platform is empty and free of debris.
2. When the display is showing LOADd CAL WT, the current setting for the calibration weight will be displayed also.
3. If the value displayed is acceptable, place the calibrated test weight on the scale platform and press the **PRINT/ENTER** key. Proceed to Step 6.
4. Otherwise, determine the exact amount of test weight to be placed on the scale platform and then using the numeric keys, enter this value. Allowable values are 1 through scale capacity (refer to the capacity label on MV2).

**NOTE:** A minimum of 50% of the scale's capacity is required. However 70% to 100% is recommended.

5. Verify that the numbers entered are the same as the amount of test weight and then press the **PRINT/ENTER** key.
6. The display will change to show PLEASE WAIT and starting at the left and proceeding right, a series of dashes will appear on the display and then starting at the right and proceeding left disappear. Next, the display will change to UNLOAD SCALE.

## UNLOAD SCALE (Unload Scale)

With the display showing UNLOAD SCALE:

1. Remove the weight from the scale platform.
2. Press the **PRINT/ENTER** key.
3. The display will change to show PLEASE WAIT and starting at the left and proceeding right, a series of dashes will appear on the display and then starting at the right and proceeding left disappear.
4. Next, the display will change to show the GRAVITY ENAbLE prompt.

# Multi-point Calibration

## CAL WEIGHT 1 (First Calibration Weight)

The display will show CAL WEIGHT 1 (which is a prompt for entry of the first calibration weight value) and placement of this amount of test weight on scale platform. Note that the MV2 defaults to half (½) of the scale capacity.

1. Make certain scale platform is empty and free of debris.
2. When the display is showing CAL WEIGHT 1, the current setting for the first calibration weight will be displayed also. If the value displayed is acceptable, place the calibrated test weight on the scale platform and then press the **PRINT/ENTER** key. Proceed to Step 5.
3. Otherwise, determine the exact amount of test weight to be placed on the scale platform and then using the numeric keys, enter this value. Allowable values are 1 through scale capacity (refer to the capacity label on MV2).
4. Verify that the numbers entered are the same as the amount of test weight and then press the **PRINT/ENTER** key.
5. The display will change to show PLEASE WAIT and starting at the left and proceeding right, a series of dashes will appear on the display and then starting at the right and proceeding left disappear. Next, the display will change to CAL WEIGHT 2.

## CAL WEIGHT 2 (Second Calibration Weight)

The display will show CAL WEIGHT 2 (which is a prompt for entry of the second calibration weight value) and placement of this amount of test weight on scale platform. Note that the MV2 defaults to full scale capacity.

1. When the display is showing CAL WEIGHT 2, the current setting for the second calibration weight will be displayed also. If the value displayed is acceptable, place the calibrated test weight on the scale platform and then press the **PRINT/ENTER** key. Proceed to Step 4.
2. Otherwise, determine the exact amount of test weight to be placed on the scale platform and then using the numeric keys, enter this value. Allowable values are 1 through scale capacity (refer to the capacity label on MV2).
3. Verify that the numbers entered are the same as the amount of test weight and then press the **PRINT/ENTER** key.
4. The display will change to show PLEASE WAIT and starting at the left and proceeding right, a series of dashes will appear on the display and then starting at the right and proceeding left disappear. Next, the display will change to UNLOAD SCALE.



**NOTE: If the same weight is used for CAL WEIGHT 1 and CAL WEIGHT 2, then the scale reverts to a Single Point calibration.**

## UNLOAD SCALE (Unload Scale)

With the display showing UNLOAD SCALE:

1. Remove the weight from the scale platform.
2. Press the **PRINT/ENTER** key.
3. The display will change to show PLEASE WAIT and starting at the left and proceeding right, a series of dashes will appear on the display and then starting at the right and proceeding left disappear.
4. Next, the display will change to show the GRAVITY ENABLE prompt.

## GRAVITY

With **GRAVITY** displayed, press the **PRINT/ENTER** key. The display will change to **GRAVITY ENAbLE**. Proceed to the Enable Gravity Constant parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the **LOCK** menu.

### **GRAVITY ENAbLE (Enable Gravity Constant)**

The Detecto MedVue® Model MV2 Weight Analyzer is equipped with a gravity constant function which means that it can be calibrated in one location and then adjusted to match the gravity at the location where it will be used.

The display will show **GRAVITY ENAbLE** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = Enable Gravity Constant and proceed to **CAL GRAVITY** parameter      no = Disable Gravity Constant and proceed to the **HOLD MODE** parameter

### **CAL GRAVITY (Calibrated Gravity Constant – Only if GRAVITY ENAbLE = YES)**

This is the gravity constant of the location where the scale was calibrated. **NOTE:** Consult the factory Tech Support for the Gravity Constant for your location.

The display will show **CAL GRAVITY** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 99,999.

### **OP GRAVITY (Operating Gravity Constant – Only if GRAVITY ENAbLE = YES)**

This is the gravity constant for the location where the scale will be operated. **NOTE:** Consult the factory Tech Support for the Gravity Constant for your location.

The display will show **OP GRAVITY** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 99,999.

## LOCK

With **LOCK** displayed, press the **PRINT/ENTER** key. The display will change to **HOLD MODE**. Proceed to the Hold Mode parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the **SYSTEM** menu.

### **HOLD MODE (Hold Mode)**

The Hold Mode Sets the locking functionality of the MV2.

The display will show **HOLD MODE** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1 or 2.

0 = Hold Mode Disabled


1 = Hold Mode Enabled. Display locked by **LOCK/RELEASE** key only

2 = Hold Mode Enabled. Display locked by **LOCK/RELEASE** key and Auto-Lock

- If **HOLD MODE = 0** or **1** is selected, proceed to the **LANGUAGE** prompt.
- If **HOLD MODE = 2** is selected, an additional prompt **HOLD BAND** will be displayed. Proceed to **HOLD BAND** (Auto Lock Reset Band) to continue setup.

With the Hold Mode enabled (**HOLD MODE = 1**), the MV2 will lock the weight shown on the display when the **LOCK/RELEASE** key is pressed and unlock the weight display when you press the key again.

If (**HOLD MODE = 2**) is selected, the MV2 will lock the weight shown on the display when the **LOCK/RELEASE** key is pressed and unlock the weight display when you press the key again as well as automatically lock the weight shown on the display after obtaining a stable weight. Note that the stable weight must exceed the **HOLD BAND** (Auto Lock Reset Band) value. Refer to the next prompt **HOLD BAND** (Auto Lock Reset Band).



### **HOLD BAND (Auto Lock Reset Band – Only if HOLD MODE = 2)**

If (**HOLD MODE = 2**) is selected, the display will change to show **HOLD BAND**. This is the number of division changes needed to reset the Auto Lock. For the Auto Lock to function, the load on the scale must rise above the **HOLD BAND** value (and remain stable). To release the Auto Lock, the load on the scale must fall below **HOLD BAND** value. At that point the next weighing operation can begin.

The display will show **HOLD BAND** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 99.

# CALIBRATION SETUP

## (Non- Protected Setup Parameters)

### SYSTEM

With SYSTEM displayed, press the **PRINT/ENTER** key. The display will change to LANGUAGE. Proceed to the Language parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the CALIBRATION menu.

### LANGUAGE (Language)

The display will show LANGUAGE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

0 = English            1 = Spanish            2 = French            3 = German



**NOTE: The default language is English. Changing the language at this prompt will change all subsequent prompts and displays to the language selected.**

### BACKLIGHT (Backlight Intensity)

The display will show BACKLIGHT and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 255.

0 = Backlight OFF            through            255 = Backlight at full intensity



**NOTE: The backlight intensity setting will affect battery life. A lower intensity setting will conserve power and maximize battery operation.**

### KEYPAD BEEP (Allow Keypad Beep)

This setting determines whether the MV2 beeps each time a key is pressed.

The display will show KEYPAD BEEP and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = The MV2 beeps each time a key is pressed.            no = MV2 does not beep when a key is pressed.

### **SLEEP MODE (Sleep Mode)**

The Sleep Mode feature also conserves power when the MV2 remains unused for a selected period of time (approximately 1 to 9 minutes). When enabled the display will show the date and time and the backlight will be turned off. The Sleep Mode feature requires the MV2 to remain at a stable zero weight to activate. Weight placed on the scale or pressing any key will activate the MV2 and return it to weight mode.

The display will show **SLEEP MODE** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 9.

**NOTE:** Selecting 0 disables the Sleep Mode feature.

### **AUTO SHUT OFF (Auto-Shut Off Feature)**

The Auto Shut Off feature will automatically turn the MV2 off (when it is not in use) after a predetermined period of inactivity (approximately 1 to 9 minutes) to conserve power. To turn the MV2 back on you must press the **ON/OFF** key.

The display will show **AUTO SHUT OFF** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 9.

**NOTE:** Selecting 0 disables the Auto Shut Off feature.

### **KEY DISABLE (Disable Key Operation)**

This parameter determines which keys (**NET/GROSS**, **TARE**, and/or **ID/HEIGHT**) are locked out during normal operation.

The display will show **KEY DISABLE** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

- 0 = No Keys Disabled
- 1 = NET/GROSS and TARE Keys Disabled
- 2 = ID/HEIGHT Keys Disabled
- 3 = NET/GROSS, TARE, and ID/HEIGHT Keys Disabled

### **ENABLE ID (ID Number Feature)**

This setting determines whether the MV2 operates in the Transaction Storage Mode (no ID storage) or in the ID Storage Mode and whether the MV2 allows an ID number to be entered prior to beginning the weighing operation. This applies to both Basic Operation and with the BMI feature enabled. **NOTE:** The ID Storage Mode allows up to 99 ID's.

The display will show **ENABLE ID** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = ID Enabled, uses ID Storage Mode

no = ID Disabled, uses Transaction Storage Mode



### **bARCOdE (Use ID Scanner)**

This setting determines whether the MV2 allows a scanner to be used to input an ID number to be entered prior to beginning the weighing operation. This applies to both Basic Operation and with the BMI feature enabled.

The display will show bARCOdE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = ID Scanner is used to input ID number      no = ID Scanner is not used

### **USE HEIGHT (Allow Height Functions)**

This setting determines whether the height functions are allowed and if the Height Units prompt is shown.

The display will show USE HEIGHT and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = The Height functions are allowed and the HEIGHT UNITS prompt will be displayed      no = The Height functions are not allowed. Proceed to the HI RESOLUTION parameter

**NOTE:** If USE HEIGHT = no (no) is selected, the time will be shown on the display in the location where the height normally would be.

### **HEIGHT UNITS (Height Units – Only if USE HEIGHT = YES and UNITS > 1)**

The display will show HEIGHT UNITS and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 or 2.

1 = Centimeters      2 = Millimeters

### **bODY MASS IdX (Body Mass Index, BMI Calculation Feature)**

This setting determines whether the MV2 performs basic weigh calculation only or performs the Body Mass Index (BMI) calculations.

The display will show bODY MASS IdX and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = Body Mass Index is Enabled and MV2 performs BMI calculations      no = Body Mass Index is Disabled and the MV2 performs basic weight calculations only

## HI RESOLUTION (High Resolution Mode)

The High Resolution Mode feature will display the weight at 10 times the standard resolution. For example, a MV2 configured for 400 lb x 0.2 lb will display weight in high resolution mode at 400 lb x 0.02 lb. **NOTE:** If `DECIMAL=3`, the High Resolution Mode is not available.

The display will show HI RESOLUTION and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting.

Allowable values are YES or no.

YES = If High Resolution Weight mode is desired

no = High Resolution Weight mode is not desired

If viewing the High Resolution Weight mode *immediately* is desired:

1. With HI RESOLUTION displayed, press **1/YES** key.
2. Press the **PRINT/ENTER** key.
3. The display will change to RS232 MODE.
4. Press the **UNITS** key twice.
5. The MV2 will reset and display the weight in High Resolution mode.

Otherwise, after selecting the value for the High Resolution mode parameter, press the **PRINT/ENTER** key to save the selection and advance to the RS232 Port Output Mode parameter to continue Calibration Setup.

## SERIAL PORT

With SERIAL PORT displayed, press the **PRINT/ENTER** key. The display will change to RS232 MODE. Proceed to the RS232 Port Output Mode parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the USB PORT menu.

### RS232 MODE (Rs232 Port Output Mode)

The display will show RS232 MODE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

- 0 = RS232 Port Output Mode is Disabled
- 1 = RS232 Port Output Mode is set for Weight On Demand
- 2 = RS232 Port Output Mode is set for Continuous Data
- 3 = RS232 Port Output Mode is set for a Printer

### BAUD RATE (Serial Output Baud Rate – Only if RS232 MODE = 1, 2 or 3)

The display will show BAUD RATE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 through 9 and 0.

- |               |                |                 |
|---------------|----------------|-----------------|
| 1 = 2400 Baud | 4 = 14.4k Baud | 7 = 38.4k Baud  |
| 2 = 4800 Baud | 5 = 19.2k Baud | 8 = 56k Baud    |
| 3 = 9600 Baud | 6 = 28.8k Baud | 9 = 57.6k Baud  |
|               |                | 0 = 115.2k Baud |

### PRINTER TYPE (Set Type Of Printer Output – Only if RS232 MODE = 3)

The display will show PRINTER TYPE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

- 0 = Print default ticket (P185 or other ASCII printer)
- 1 = Print journal line
- 2 = Print label (P220)
- 3 = Print nControl ticket (must have loaded a ticket format, otherwise will use default ticket)

### END OF PRINT (End Of Print Linefeeds – Only if RS232 MODE = 3)

At the end of data sent to a printer, the MV2 can send a number of carriage return/line feeds to space the paper in the printer to a desired position for withdrawal or for the next print.

The display will show END OF PRINT and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 9.

## USB PORT

With USB PORT displayed, press the **PRINT/ENTER** key. The display will change to USB MOD. Proceed to the USB Port Output Mode parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the TIME SETTINGS menu.

**NOTE:** If an Ethernet or WiFi option card is installed, pressing the **ID/HEIGHT** key will advance to the ETHERNET/WIFI menu.

### USB MOD (USB Port Output Mode)

The display will show USB MOD and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2, 3 or 4.

- 0 = USB Port Output Mode is Disabled
- 1 = USB Port Output Mode is set for Weight On Demand  
(requires serial driver .inf file for Windows)
- 2 = USB Port Output Mode is set for Continuous Data  
(requires serial driver .inf file for Windows)
- 3 = USB Port Output Mode is set for a Printer  
(only works for printers with an internal font)
- 4 = USB Port Output Mode is set for Welch Allyn Operation  
(Requires device to have an activated license. Must purchase Detecto WACONNECT)

### PRINTER TYPE (Set Type Of Printer Output – Only if USB MOD = 3)

The display will show PRINTER TYPE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

- 0 = Print default ticket
- 1 = Print journal line
- 2 = Print label
- 3 = Print nControl ticket (must have loaded a ticket format, otherwise will use default ticket)

### END OF PRINT (End Of Print Linefeeds – Only if USB MOD = 3)

At the end of data sent to a printer, the MV2 can send a number of carriage return/line feeds to space the paper in the printer to a desired position for withdrawal or for the next print.

The display will show END OF PRINT and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 9.

## **ETHERNET/WIFI – (If Option Card Installed)**

With ETHERNET/WIFI displayed, press the **PRINT/ENTER** key. The display will change to ETHERNET MODE. Proceed to the Ethernet Port Output Mode parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the TIME SETTINGS menu.

### **ETHERNET MODE (Ethernet Port Output Mode)**

The display will show ETHERNET MODE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

- 0 = Ethernet Port Output Mode is Disabled
- 1 = Ethernet Port Output Mode is set for Weight On Demand
- 2 = Ethernet Port Output Mode is set for Continuous Data
- 3 = Ethernet Port Output Mode is set for a Printer (not for connection to a printer but can output print data to PC if connected)

### **PRINTER TYPE (Set Type Of Printer Output – Only if ETHERNET MODE = 3)**

The display will show PRINTER TYPE and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0, 1, 2 or 3.

- 0 = Print default ticket
- 1 = Print journal line
- 2 = Print label
- 3 = Print nControl ticket (must have loaded a ticket format, otherwise will use default ticket)

### **END OF PRINT (End Of Print Linefeeds – Only if ETHERNET MODE = 3)**

At the end of data sent to a printer, the MV2 can send a number of carriage return/line feeds to space the paper in the printer to a desired position for withdrawal or for the next print.

The display will show End OF PRINT and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 9.

### **RESET ETHERNET (Reset Ethernet Settings to Factory Default Settings)**

The display will show RESET ETHERNET and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

- YES = Ethernet settings are reset to factory default settings
- no = Ethernet settings are not reset to factory default settings

## TIME SETTINGS

With **TIME SETTINGS** displayed, press the **PRINT/ENTER** key. The display will change to **YEAR [YYYY]**. Proceed to the **YEAR** parameter.

Otherwise, press the **ID/HEIGHT** key to advance to the **SCALE** menu.

### **YEAR [YYYY] (Year)**

The display will show **YEAR [YYYY]** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 2012 through 2099.

### **MONTH [ 1- 12] (Month)**

The display will show **MONTH [ 1- 12]** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 through 12.

1 = January	5 = May	9 = September
2 = February	6 = June	10 = October
3 = March	7 = July	11 = November
4 = April	8 = August	12 = December

### **DAY [ 1-3 1] (Day)**

The display will show **DAY [ 1-3 1]** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 1 through 31.

### **HOURL [0-23] (Hour)**

The display will show **HOURL [0-23]** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 23.

**NOTE:** The hour is entered in a 24-hour format. When entering the hour after noon (12:00 PM), you must add 12 to time. For example, 3:00 PM would be entered as 15.

### **MINUTE [0-59] (Minutes)**

The display will show **MINUTE [0-59]** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 59.

### **SECONd [0-59] (Seconds)**

The display will show **SECONd [0-59]** and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise, using the numeric keys, enter a new value and then press the **PRINT/ENTER** key to save it.

Allowable values are 0 through 59.

## **24 HOUR TIME (Use 24 Hour Time Format)**

The display will show 24 HOUR TIME and the current setting. If the value displayed is acceptable, press the **PRINT/ENTER** key to save it. Otherwise use the numeric keys **0/NO** or **1/YES** and enter a new setting. Press the **PRINT/ENTER** key to save it.

Allowable values are YES or no.

YES = Use 24 Hour Time Format  
(3PM displays 15:00)

no = Use 12 Hour Time Format  
(3PM displays 3:00)

## **Calibration Setup is Complete**

The Calibration Setup process has been completed. The MV2 will reset and then display weight. Remove power from the MV2, re-install the end cap and secure with the two (2) screws removed earlier. The MV2 is now ready for normal operation.





# SETUP REVIEW


## (Non-Protected Setup Parameters)

The MedVue Model MV2 Medical Weight Analyzer allows several operational parameters to be reviewed and changed as necessary without having to remove the end cap and pressing the calibration switch to enter the Calibration Setup mode.

### To Enter Setup Review

With the MV2 on and in the Gross weight mode:

1. Press and hold the **CLEAR** ( ← ) key for approximately 3 seconds.
2. The display will change to show **SETUP REVIEW**.
3. Release the **CLEAR** ( ← ) key.
4. The display changes to show:



Refer to the instructions listed in the Calibration Setup (Non- Protected Setup Parameters) section (page 33) of this manual for information on how to change these parameters.

Parameters in Setup Review will be processed in the following sequence:

#### SYSTEM

LANGUAGE	Select MV2 operation language.
BACKLIGHT	Set Backlight Intensity
KEYPAD BEEP	Enable or Disable MV2 emitting a beep when a key is pressed.
SLEEP MODE	Disable or select number of minutes of inactivity for sleep mode
AUTO SHUT OFF	Disable or select number of minutes for automatic shutoff timer
KEY DISABLE	Enable or Disable <b>NET/GROSS</b> , <b>TARE</b> and/or <b>ID/HEIGHT</b> keys
ENABLE ID	Enable or Disable transaction storage and ID number input feature
BARCODE	Enable or Disable barcode scanning of ID's
USE HEIGHT	Enable or Disable Height Functions
HEIGHT UNITS	Select Centimeters or Millimeters for height measuring units
BODY MASS IDX	Enable or Disable Body Mass Index (BMI) calculation feature
HI RESOLUTION	Enable High Resolution Weight Mode feature

## SERIAL PORT

RS232 MODE	Set the mode of the RS232 serial port
BAUD RATE	Set baud rate for RS232 serial port (If RS232 MODE = 1, 2 or 3)
PRINTER TYPE	Set the type of printer output (If RS232 MODE = 3)
End OF PRINT	Set Number of End Of Print linefeeds (If RS232 MODE = 3)

## USB PORT

USB MODE	Set the mode of the USB port
PRINTER TYPE	Set the type of printer output (If USB MODE = 3)
End OF PRINT	Set Number of End Of Print linefeeds (If USB MODE = 3)

## ETHERNET/WIFI *If option card is present*

ETHERNET MODE	Set the mode of the Ethernet port
PRINTER TYPE	Set the type of printer output (If ETHERNET MODE = 3)
End OF PRINT	Set Number of End Of Print linefeeds (If ETHERNET MODE = 3)
RESET ETHERNET	Reset Ethernet settings to factory default settings


## TIME SETTINGS

YYYY [YYYY]	Set the Year
MONTH [ 1- 12]	Set the Month of the year
DAY [ 1-30]	Set the Day of the month
HOUR [0-23]	Set the Hours portion of the time (Entered in a 24-hour format)
MINUTE [0-59]	Set the Minutes portion of the time
SECOND [0-59]	Set the Seconds portion of the time
24 HOUR TIME	Select to use 24 hour time format (3PM displays 15:00)

### **Setup Review Is Complete**

The setup review process has been completed. The MV2 will reset and then display weight. The MV2 is now ready for normal operation.

## ERROR AND STATUS DISPLAYS

Display	Meaning
ERROR	General error, invalid keypad entry was attempted.
ERR OVERFLOW	Attempting to display a negative number greater than -9,999 or a positive number greater than 99,999
ERR ZERO LIM	Indicates an attempt to zero a weight outside scale zero range. Refer to the Calibration Setup <b>ZERO</b> Menu, <b>ZERO LIMIT</b> (Two Percent Zero Tracking Range Limit) parameter.
ERR UNSTABLE	Motion is present when MV2 is attempting to perform one of the following operations: Power Up Zero or Zero Weight Display
OVER CAPACITY	Scale weight exceeds scale capacity.
	When the batteries are near the point they need to be recharged, the battery status symbol <u>will not show</u> any bars. If the battery voltage drops too low for accurate weighing, the MV2 will automatically shut off and you will be unable to turn it back on.
CAL REQUIRED	Indicates improper stored calibration data. Calibration of MV2 is required. Consult your scale service representative.
ERR ANALOG	An analog to digital convertor communication error has been detected. Consult your scale service representative.
ANALOG LOW	The load cell input is <u>below</u> the range of the MV2. Consult your scale service representative.
ANALOG HI	The load cell input is <u>above</u> the range of the MV2. Consult your scale service representative.
ERR EEPROM	NOVRAM failure. Consult your scale service representative.

## BEFORE YOU CALL FOR SERVICE

Problem	Possible Solutions
Display does not turn on	<p>AC Operation:</p> <ul style="list-style-type: none"> <li>• Is AC power supply fully inserted into wall receptacle?</li> <li>• Check wall receptacle for proper AC power. Try another electrical appliance in same receptacle, does it work?</li> <li>• Check circuit breaker.</li> <li>• Has there been power failure?</li> </ul> <p>Battery operation:</p> <ul style="list-style-type: none"> <li>• Check if batteries are installed and correctly.</li> <li>• Is the battery status symbol empty (no bars are displayed)?            Connect the optional 12 VDC AC power adapter's connector into the jack on the back of the MV2 and then plug the power adapter into the proper electrical outlet to operate the MV2 and recharge the batteries.            Otherwise, remove discharged batteries and replace with fully charged ones and place discharged batteries in an external charger to recharge.</li> </ul>
Incorrect weight is displayed	Insure that scale platform isn't touching an adjacent object. Have proper operation procedures been followed?
Weight is not displayed	Refer to Error and Status Display section.
Height is not displayed	The digital height rod was not in its "starting position" before turning on the MV2. The inner sliding tube <u>must</u> be completely down inside outer stationary tube and headpiece folded flat against stationary tube prior to turning on the MV2.

## CARE AND CLEANING



**NOTE:** The MV2 contains no user-serviceable parts and maintenance should be limited to an occasional cleaning and battery replacement as required.


- Do not submerge the MV2 in water, pour or spray water directly on it to clean. The MV2 is not waterproof and covering it with water will damage it and void the warranty.
- Always remove power before cleaning.
- Do not use wire brushes, abrasives, or cleaning tools such as steel pads and scrapers, which will scratch the painted surface. Instead, use soft cloths or plastic scouring pads for cleaning.
- When possible use treated water. Hard water can leave behind deposits. Soft water is much gentler on the painted steel's surface.
- Avoid the use of acetone, thinner or other volatile solvents and abrasive type cleaners for cleaning. If required, a mild solvent such as mineral spirits can be used to remove oil, grease, tars, wax, and similar substances. Use a cloth dampened with mineral spirits and apply only to areas that are contaminated. Follow up the use of this mild solvent with detergent cleaning and rinsing.

## SECURITY SEAL INSTALLATION

If your metrology laws require the use of physical sealing, a lead and wire security seal can be installed on the MV2 to prevent the end cap from being removed from it to gain access to the calibration switch thereby preventing unauthorized access to the calibration adjustments.

To install a lead and wire security seal, refer to Figure No. 13 and follow the steps provided below:

1. Make certain the two drilled screws securing the left end cap (as viewed from the MV2 front) are securely tightened before proceeding.
2. Thread the sealing wire through the hole in the drilled screws.
3. Pull the wire tight and install the lead seal.
4. The screws can not be removed without damaging the seal.




**Figure No. 13**




## PARTS IDENTIFICATION

Item	Qty	Part Number	Description
1	1	5930-B126-08	LABEL: ETL
2	1	593GR986	SERIAL TAG ASSY
3*	2	6021-2071	SCW FILLISTER. MACHINE-SCW 06-32X.250
4	1	6021-1065	SCW THUMB..... MACHINE-SCW 06-32X.375
5	7	6021-2069	SCW TRUSS HEAD SHEET METAL #6 X .50
6	2	6021-6008	SCW TRUSS HEAD SHEET-METAL #6 X.375
7	2	6024-0126	WASHER FLAT FOR 5/32" POP RIVET
8	1	6560-1125	ADHESIVE LOCTITE 290 THRD LCKR, WICKING
10	4	6610-2000	JACK SOCKET, 1 EA (NO WASHERS OR NUTS)
11	1	6610-5119	BATTERY HOLDER 6-C CELLS
12	7	6680-0004	WASHER LOCK INT TOOTH #6 TYPE A Z-PL
13	4	6680-0052	WASHER LOCK HELICAL #4 REG Z-PL
14	2	6680-0214	RIVET POP, BH AL 5/32 DIA x .126-.187 GR
15	1	8555-0526-0A	BACK PANEL WELDMENT
16	1	8555-0528-08	FRONT PANEL
17	1	8555-0529-08	BATTERY COVER
18	1	8555-B461-0A	CABLE: BATTERY
20	1	8555-B510-08	RUBBER BALL
21	2	8555-C213-08	END CAP
23	1	8555-D501-0A	PCB ASS'Y MEDVUE CONTROLLER
24	1	8555-D502-08	KEYPAD : MEDVUE
**	1	MV2PWR	OPTIONAL AC ADAPTER 100-240VAC/12VDC @ 1 AMP

\* INSTALLED ON RIGHT END CAP (AS VIEWED FROM MV2 BACK)

\*\* NOT SHOWN







# APPENDIX A

## (Top Plate Mounting Configurations)

**“Type One”**  
**(0033-B104-0A)**  
Uses Center Mounting Hole



**“Type Three”**  
**(0033-B104-0A)**  
Uses Front Mounting Hole





### Models using “Type One”

6495  
6437  
6437DHR  
6437KGDHR  
6438  
6439  
6439KG  
6439M  
6447  
6449  
6449M

### Models using “Type Three”



8430  
8430M  
8431  
8431M  
8437

**“Type Four”  
(8555-B517-08)  
TRANSITION PLATE**





**Models using “Type Four”**  
6550

**“Type Five”  
(1930-C199-08)  
MOUNTING BASE**



**Models using “Type Five”**  
6854DHR  
6857DHR  
6857KGDHR

**“Type Desk/Wall Mount”  
(8555-B516-08)  
DESKTOP PLATE**



**Models using “Desk/Wall Mount”**

- 6500
- BRW1000
- CD-1000D
- CR-500D
- FHD-133-II
- FHD-143-II
- FHD-144-II
- FHD-164-II
- GP-400-MV2
- GP-600-MV2
- GP-600-RP

**WALL MOUNTING**

Mount using two (2) #10 screws placed 2.00 inches (51 mm) apart in the wall.



## APPENDIX B

### (Connecting to a Welch Allyn CVSM Device)

This section describes the required settings of the MedVue Model MV2 Medical Weight Analyzer in order to connect to a Welch Allyn CVSM device.

To enter Setup on the MedVue Model MV2 Medical Weight Analyzer, remove the two (2) Phillips head screws from the right end cap (as viewed from the back of the MV2) and remove the end cap. Look for a small hole under the end cap with a red switch button behind it. It can be found near the bottom edge of the Model MV2 Medical Weight Analyzer. Refer to the CALIBRATION SETUP (Protected Setup Parameters) section for more information.

1. Press the switch with a paper clip or similar tool and you should see **SEtUP** on the top of the display and **SCALe** at the bottom.
2. If it is desired for the MedVue to automatically send the weight, height, and BMI when the weight is locked, the first setting that must be changed is the automatic lock feature of the MedVue. While in setup, press the **ID/H** key until the display shows **LOCK** at the bottom.
3. Press the **ENTER** key and you should see **HOLD MODE** at the bottom of the MedVue display. To enable the auto locking feature, change this setting to "2" and then press the **ENTER** key.
4. The next setting that appears is the **HOLD BAND**. Enter "10" for this setting and then press the **ENTER** key. Press the **UNITS** key to return to the setup menu.
5. The next setting that must be correct in order for communication to work is the **USB MODE**. While in the setup menu, press the **ID/H** key until **USB PORT** is displayed at the bottom of the MedVue display and then press the **ENTER** key.
6. Change the **USB MODE** to "4" and then press the **ENTER** key.

**You should now have all the correct settings in order to communicate with the Welch Allyn device.**

#### Auto-Lock Operation

If you have the MedVue set to auto-lock, simply apply weight to the scale. When the scale settles and the weight locks, the weight, height and BMI information will be transmitted.

The height and BMI do not lock with the weight. If the height was not entered in prior to applying weight on the scale, only the weight will be transmitted to the Welch Allyn device.

If the operator enters in the height, either by using the digital height rod or entering the height using the keypad, the height and calculated BMI information will be transmitted at that time.

#### Manual Operation

If you do not have the auto-lock enabled, pressing the **PRINT/ENTER** key on the MedVue keypad will transmit the weight, height and BMI information.

Pressing the **PRINT/ENTER** key will clear the current height and BMI. If it becomes necessary to retransmit the information, it will be necessary to enter the height again.

## STATEMENT OF LIMITED WARRANTY

Detecto Scale warrants its equipment to be free from defects in material and workmanship as follows: Detecto warrants to the original purchaser only that it will repair or replace any part of equipment which is defective in material or workmanship for a period of two **(2) years from date of shipment**. Detecto shall be the sole judge of what constitutes a defect.

During the **first ninety (90) days** Detecto may choose to replace the product at no charge to the buyer upon inspection of the returned item.

**After the first ninety (90) days**, upon inspection of the returned item, Detecto will repair or replace it with a remanufactured product. The customer is responsible for paying for the freight both ways.

This warranty does not apply to peripheral equipment not manufactured by Detecto; this equipment will be covered by certain manufacturer's warranty only.

This warranty does not include replacement of expendable or consumable parts. This does not apply to any item which has deteriorated or damaged due to wear, accident, misuse, abuse, improper line voltage, overloading, theft, lightning, fire, water or acts of God, or due to extended storage or exposure while in purchaser's possession. This warranty does not apply to maintenance service. Purchased parts will have a ninety (90) day repair or replacement warranty only.

Detecto may require the suspect product to be returned to the factory; item(s) must be properly packed and shipping charges prepaid. A return authorization number must be obtained for all returns and marked on the outside of all returned packages. Detecto accepts no responsibility for loss or damage in transit.

# STATEMENT OF LIMITED WARRANTY

## Conditions Which Void Limited Warranty

This warranty shall not apply to equipment which:

- A.) Has been tampered with, defaced, mishandled or has had repairs and modifications not authorized by Detecto.
- B.) Has had serial number altered, defaced, or removed.
- C.) Has not been grounded according to Detecto's recommended procedure.

## Freight Carrier Damage

Claims for equipment damaged in transit must be referred to the freight carrier in accordance with freight carrier regulations.

This warranty sets forth the extent of our liability for breach of any warranty or deficiency in connection with the sale or use of the product. Detecto will not be liable for consequential damages of any nature, including but not limited to, loss of profit, delays or expenses, whether based on tort or contract. Detecto reserves the right to incorporate improvements in material and design without notice and is not obligated to incorporate improvements in equipment previously manufactured.

The foregoing is in lieu of all other warranties, express or implied including any warranty that extends beyond the description of the product including any warranty of merchantability or fitness for a particular purpose. This warranty covers only those Detecto products installed in the forty-eight (48) contiguous continental United States.



Ph. (800) 641-2008  
E-mail: [detecto@cardet.com](mailto:detecto@cardet.com)  
203 E. Daugherty  
Webb City, MO 64870

08/27/2014  
Printed in USA  
D268-WARRANTY-DET-A



Printed in USA

8555-0530-0M Rev D 10/14



203 E. Daugherty, Webb City, MO 64870 USA

Ph: 417-673-4631 or 1-800-641-2008

Fax: 417-673-2153

Technical Support: 1-866-254-8261

E-mail: [tech@cardet.com](mailto:tech@cardet.com)